

Perennial & Annual Ryegrass 919 Moisture Meter Chart

Instructions

- Sample size 175 grams
- Calibrate meter to 63
- Chart does not compensate for temperature. Seed temperature must be Temp 19°C - 25°C (66°F - 78°F) to assess moisture accurately.
- Dry is 12%
- If sample fills cell above test cone
- Sample 150 grams
- Calibrate at 73

Dial Reading	% Moisture	Dial Reading	% Moisture
1	9.14	21	13.62
2	9.36	22	13.85
3	9.59	23	14.07
4	9.81	24	14.29
5	10.03	25	14.51
6	10.26	26	14.74
7	10.48	27	14.96
8	10.71	28	15.19
9	10.93	29	15.41
10	11.15	30	15.53
11	11.38	31	15.86
12	11.50	32	16.08
13	11.83	33	16.31
14	12.05	34	16.53
15	12.27	35	16.75
16	12.50	36	16.98
17	12.72	37	17.20
18	12.95	38	17.43
19	13.17	39	17.65
20	13.39	40	17.87

 Dry